

PARASURE

Bowie and Dick Test

121C for 8 minutes

Conforms to EN ISO 11140–4
CE Mark
GTIN numbered

Independently validated to mirror the performance of the original
Towel pack

Diagnostic capabilities:
- air removal
- wet steam
- super heated steam or presence of non-condensable gases

Clear and accurate colour change

Easy to interpret – easy to record

Non-toxic, single use, disposable



Bowie and Dick Daily Steriliser Test (B&D Test) The ParaSure B&D test pack fulfils the requirements of both EN 285 and EN ISO 17665

- The test is performed at the beginning of each working shift in every

vacuum assisted steam steriliser.

- This non-toxic test is not a test of sterility but a test of the mechanical

function of the steriliser to ensure rapid and even steam penetration into a small porous load.

Order Details

PS1030R

Bowie and Dick Test pack Box of 30



WWW.PARASUREORGANICS.COM

Registered Address:

ParaSure Limited Units B10/11 Oakendene Estate Bolney Rd Cowfold West Sussex RH13 8AZ England



PARASURE

Daily Bowie & Dick Test Pack

TECHNIC/	L SPECIFICATIONS
Name	Daily Bowie & Dick Test Pack, steam penetration
Classification:	EN-ISO 11140-1 (Type 2), EN-ISO 11140-
Operating temp:	121°C
Time:	2-11 min
Storage conditions:	10°C - 30°C
RH:	30 - 70%
Lot no.	yes
Prod date:	yes
Exp date:	3 years
Correct symbols:	in accordance with EN-ISO 11140-4
Indicator:	both internal and external, external = process indicator, internal = diagnostic indicator
Disp/single use:	yes
Information on pack: Cover Filtration medium	Wrapper barrier system
Extra information: Sheet Coated Non-toxic/Lead free Lot no.	silk screen printed yes yes yes
Recordable Information: Hospital Department Autoclave no Cycle no Date Operator Supervisor Pass/fail result Colour change	yes yes yes yes yes yes yes yes pes yes

For further information about any of our products, please visit: www.parasure.co.uk



